

## Network Notification – Humana Healthy Horizons in Ohio

**To:** Humana Healthy Horizons® in Ohio provider network  
**From:** Humana Healthy Horizons in Ohio  
**Subject:** Provider network notification – FDA MedWatch - HAMILTON-C1, T1, MR-1 Ventilators Recalled for Leaks and Short Circuits  
**Effective Date:** October 20, 2023

---

### **Hamilton Medical Inc. Recalls HAMILTON-C1, T1, MR-1 Ventilators for Capacitor Leaks and Short Circuits**

Hamilton Medical, Inc. is recalling the **HAMILTON-C1, T1, MR-1 ventilators** because the capacitors may leak electrolyte fluid onto the ventilator’s control board. If the control board contacts the electrolyte fluid, the control board or installed spare parts could short circuit. As a result of the short circuit, the ventilator may switch to “Ambient State.”

The FDA has identified this as a **Class I recall**, the most serious type of recall. Use of these devices may cause serious injuries or death.

### **Questions?**

If you have questions about this recall, contact Hamilton Medical Inc. at 800-426-6331 or email, [reno.techsupport@hamiltonmedical.com](mailto:reno.techsupport@hamiltonmedical.com).